



DECLARATION OF CONFORMITY

Manufacturer's Name: Hill-Rom, Inc.

Manufacturer's Address: 1069 State Route 46 East, Batesville, Indiana, USA 47006



Name of Device: Affinity® Four Birthing Bed

Model Numbers: P3700 including features and accessories listed in the Instructions For Use³
P3700 (beds will be marked with Model No. option code P3700@XXXXXX¹)

Serial Numbers: B277AA0040² (first serial no. on Oct. 02, 2000) through present.

We hereby declare that the above mentioned medical devices comply with the provisions of the European Medical Devices Directive 93/42/EEC. The Affinity® Four Birthing Bed is a Class I device, per Annex IX, Rule 12. This declaration is based on conformity with Annex VII.

European Union Representative: **Hill-Rom SAS**
B.P. 14 - Z.I. du Talhouet
56330 Pluvigner, France

Signature:		
Name & Title:	Lawrence E. Luff Codes & Standards Engineer, Lead (prepared by)	Ann Waterhouse Director Regulatory Affairs (approved by)
Date:	<u>6 August 2013</u>	<u>7 Aug 2013</u>

¹ P3700@XXXXXX where P3700 = Affinity; @ = bed rev letter (A, B, C, D...) which will roll over for major model changes ; The six 'X's represent a number which corresponds to a unique mix of options and features (such as MAINS input voltage, power cord type, labels (Ger, Fr, etc.)

² Y DD PPCCCC where Y = year (A = 1999, B = 2000,...H=2012); DDD = day of year; PP = Product Line (AA = Affinity); CCCC = Continuous Counter which rolls over (0001-9999).
 EXAMPLE, such as B277AA0040 where: B= year 2000, 277 = 277 day of year, AA=Affinity & 0040 = count

³ Availability of specific features and accessories, including mattresses, may be limited by specific country requirements.